

TECNO 25



Suction Unit

Model

Typology (MDD 93/42/EEC)

Classification UNI EN ISO 10079-1

Main Voltage

Power consumption

Fuse

Maximum suction aspiration (without jar)

Minimum suction aspiration (without jar)

Maximum flow (without jar)

Weight

Battery Holding Time

Battery Time Charge

Dimension

Functioning

Working condition

Jar

TECNO 25

Medical Device Class IIa

HIGH VACUUM / HIGH FLOW

230 V ~ / 50 Hz

110 VA

F 1 x 1.6A L 250 V

-80kPa (- 0.80bar)

Less than -40kPa (-0.40 bar)

40 l/min

3.6 Kg

-

-

350 x 210 x 180 mm

NON-STOP OPERATED

Room temperature: 5 ÷ 35 °C

Room humidity percentage: 10 ÷ 93 % RH

Atmospheric pressure: 700 ÷ 1060 hPa

1lt,2lt

